

Combustible (LEL) Sensor Correction Factor

SemeaTech combustible (LEL) sensors are capable of detecting a wide range of flammable gases. These sensors are sensitive to various hydrocarbons, including methane, propane, and other organic vapors through oxidation. When a flammable gas contacts the heated catalyst, it combusts internally, causing a temperature increase that alters the resistance of the active element and then it is converted into an electrical signal. The strength of signal is proportional to the concentration of the gas.

LEL sensors require a certain amount of oxygen to function properly, typically at least 10-15% Vol in the air. If the oxygen concentration drops below this level, the sensor's ability to detect flammable gases significantly diminishes. This is because the combustion process is hindered, leading to incomplete oxidation and a lower sensor response. Conversely, excessively high oxygen concentrations may cause false alarms.

Methane sensitivity and the correction factor of LEL sensors may change over time due to various factors related to sensor aging and usage. The primary factor is sensor degradation. The components of the sensor, including the catalyst used for methane detection, can degrade due to exposure to environmental factors, contaminants, or prolonged use, which affect performance and accuracy. This can also result from calibration drift and changes in temperature and humidity.

LEL sensors are typically calibrated with a specific gas, such as methane, but they are capable of detecting various combustible gases. Correction factors are applied to ensure accurate measurements across different gases. These factors adjust the sensor's readings to reflect the actual concentration of the detected gas.

Correction factors are used in various calibration scenarios for LEL sensors to ensure accurate %LEL measurements of different gases:

Methane is the most common reference gas to do the LEL sensors calibration. When calibrating with methane, multiply the reading by the Correction Factor (CF Value) to obtain the actual %LEL of the measured gas. For corrected span gas concentration, calibrate the unit using methane and input the equivalent "corrected" span gas concentration when prompted. For example, to measure in Ethanol LEL units, multiply the percentage of methane LEL by 1.7 (CF Value of Ethanol). For calibration by non-methane compounds, calculate the value by dividing the CF value of the calibration compound by that of the span compound. For instance, on the Isopropanol scale, divide the table value by 2.6 (Isopropanol CF value), resulting in a new value of $2.8/2.6 = 1.08$ for Ethyl benzene.

Chemical	100% LEL		Chemical	100%LEL	
	(Vol%)	LEL CF		(Vol%)	LEL CF
Acetaldehyde	4.0	1.8	Dimethyl sulfide	2.2	2.3
Acetic acid	4.0	3.4	Dioxane, 1,4-	2.0	2.5
Acetic Anhydride	2.7	2.0	Ethane	3.0	1.4
Acetone	2.5	2.2	Ethanol	3.3	1.7
Acetylene	2.5	2.8	Ethene	2.7	1.4
Allyl Alcohol	2.5	1.7	Ethyl acetate	2.0	2.2
Ammonia	15.0	0.8	Ethylamine	3.5	1.4
Aniline	1.3	3.0	Ethyl benzene	0.8	2.8
Benzene	1.2	2.2	Ethyl bromide	6.8	0.9
Butadiene, 1,3-	2.0	2.5	Ethyl chloride	3.8	1.7
Butane, n-	1.9	2.0	Ethyl ether	1.9	2.3
Butane, i-	1.8	1.8	Ethyl formate	2.8	2.4
Butanol, n-	1.4	3.0	Ethyl mercaptan	2.8	1.8
Butanol, i-	1.7	2.5	Ethyl methyl ether	2.0	2.3
Butanol, t-	2.4	1.8	Ethyl pentane	1.2	2.4
Butene-1	1.6	2.1	Ethylene oxide	3.0	2.3
Butene-2, cis	1.7	2.1	Gasoline,	1.3	2.1
Butene-2, trans	1.8	1.9	Heptane, n-	1.1	2.4
Butyric acid	2.0	2.4	Hexadiene, 1,4-	2.0	1.5
Carbon disulfide	1.3	/	Hexane, n-	1.1	2.3
Carbon monoxide	12.5	1.2	Hydrazine	2.9	2.1
Carbonyl sulfide	12.0	1.0	Hydrogen	4.0	1.1
Chlorobenzene	1.3	3.0	Hydrogen cyanide	5.6	2.0
Chloropropane, 1-	2.6	1.8	Hydrogen sulfide	4.0	/
Cyanogen	6.6	1.1	Isobutene (Isobutylene)	1.8	1.5
Cyclohexane	1.3	2.5	Isopropanol	2.0	2.6
Cyclopropane	2.4	1.5	Methane	5.0	1.0
Decane, n-	0.8	3.4	Methanol	6.0	1.5
Dichloroethane, 1,2-	6.2	1.5	Methyl acetate	3.1	2.2
Dichloromethane	13.0	1.0	Methylamine	4.9	1.3
Dimethylbutane	1.2	2.7	Methyl bromide	10.0	1.1
Dimethylpentane, 2,3-	1.1	2.3	Methyl chloride	8.1	1.3

CF Value Table

Chemical	100% LEL		Chemical	100%LEL	
	(Vol%)	LEL CF		(Vol%)	LEL CF
Methylcyclohexane	1.2	2.6	Pentane, n-	1.5	2.2
Methyl ether	3.4	1.7	Pentane, i-	1.4	2.3
Methyl ethyl ketone	1.4	2.6	Pentane, Neo-	1.4	2.5
Methyl formate	4.5	1.9	Pentene, 1-	1.5	2.3
Methyl hexane	1.2	2.4	Phosphine	1.6	0.3
Methyl mercaptan	3.9	1.6	Propane	2.1	1.6
Methylpentane	1.2	2.7	Propanol, n-	2.2	2.0
Methyl propionate	2.5	2.1	Propene	2.0	1.5
Methyl n-propyl ketone (2-pentanone)	1.5	2.7	Propylamine, n-	2.0	2.1
Naphthalene	0.9	2.9	Propylene oxide	2.3	2.6
Nitromethane	7.3	2.1	Propyl ether, iso-	1.4	2.3
Nonane, n-	0.8	3.2	Propyne	1.7	2.3
Octane, n-	1.0	2.9	Toluene	1.1	2.6
Pentane, n-	1.5	2.2	Triethylamine	1.2	2.5
Pentane, i-	1.4	2.3	Trimethylamine	2.0	1.9
Pentane, Neo-	1.4	2.5	Trimethylbutane	1.2	2.3
Pentene, 1-	1.5	2.3	Turpentine	0.8	2.9
Phosphine	1.6	0.3	Vinyl chloride	3.6	1.8
Propane	2.1	1.6	Xylene, m-	1.1	2.7
Propanol, n-	2.2	2.0	Xylene, o-	0.9	3.0
Propene	2.0	1.5	Xylene, p-	1.1	2.8

CF Value Table

Caution:

Several gases and substances may cause LEL sensors to experience baseline shifts and sensitivity loss:

1. Silicones: Catalyst in LEL sensors will be poisoned when exposure to silicone compounds.
2. Sulfur Compounds: Mainly Hydrogen Sulfide (H₂S) will poison the catalyst to decrease the sensitivity and lead to baseline shifts.